

Certificate

Efficiency of Particle Retention

ultrex® P-GS

Pore size: 5 micron nominal

| particle size micron | particles counts per minute before filter | particles counts per minute after filter | efficiency % air | efficiency % steam |
|-------------------------|--|---|------------------------|--------------------------|
| 0,5 | 7288 | 1337 | 82 | 78 |
| 1 | 8244 | 3 | 99,96 | 96 |
| 2 | 20250 | 1 | 99,995 | 97 |
| 5 | 1736 | 0 | 100 | 98 |

ultrafilter international AG



Michael Schuster
Technical Director